



NATIONAL TESTING LABORATORIES, LTD.
 556 South Mansfield, Ypsilanti, MI 48197-5166
 (440) 449-2525 • Fax (440) 449-8585



ANALYTICAL REPORT

Date: 07/13/06 Report #: 655357 Laboratory ID #: 0055
 PAGE 1 OF 2 SAMPLE CODE: 655357

Client: RUSH BEVERAGES INC.
 PO BOX 207
 BLUE ISLAND, IL 60406-

Date Collected: SEE OPENING INFO
 Time Collected: SEE OPENING INFO
 SOURCE: MOUNTAIN SPRING WATER
 E WATER
 FINISHED PRODUCT
 16.9 OZ/ PROD CODE: EXP 051908

Date received at lab: 06/23/06 Time received at lab: 10:00
 Collected by : R.CORR

OPENING INFORMATION: METALS AND METHOD 300.0 OPENED 06/23/06 @ 10:45 BY
 JJL, AND ALL OTHER INORGANIC CHEMICALS & PHYSICAL FACTORS OPENED 06/27/06
 @ 12:10 BY PC.

The results herein conform to NELAC standards, where applicable, unless
 otherwise narrated in the body of the report. Test results were
 calculated on a wet weight basis. The uncertainty of the test results
 are available upon request.

PLEASE NOTE: The Reporting Limits have been changed due to sample matrix
 interference for Chromium to 0.005 mg/L, and for Selenium to 0.008 mg/L.

NOTE: "*" The MCL (Maximum Contaminant Level) or an established
 guideline has been exceeded for this contaminant.

"ND" This contaminant was not detected at or above our lower
 reporting limit (LRL). "NA" Not Analyzed

Fed Analysis Performed Id #	Method	MCL (mg/l)	LRL	Level Detected	Anal Date

Inorganic chemicals - metals:					
1002 Aluminum	200.7	0.2**	0.05	ND	07/07/06
1074 Antimony	200.8	0.006	0.001	ND	07/07/06
1005 Arsenic	200.8	0.010	0.002	ND	07/07/06
1010 Barium	200.8	2	0.10	ND	07/07/06
1075 Beryllium	200.8	0.004	0.001	ND	07/10/06
1079 Boron	200.7	---	0.10	ND	07/07/06
1015 Cadmium	200.8	0.005	0.001	ND	07/07/06
1016 Calcium	200.7	---	2.0	14	07/07/06
1020 Chromium	200.8	0.1	0.005	ND	07/07/06
1022 Copper	200.8	1.3**	0.002	0.002	07/07/06
1028 Iron	200.7	0.3**	0.020	ND	07/07/06
1030 Lead	200.8	0.015	0.001	ND	07/07/06
1031 Magnesium	200.7	---	0.10	35	07/07/06
1032 Manganese	200.8	0.05**	0.004	ND	07/07/06
1035 Mercury	200.8	0.002	0.0002	ND	07/07/06
1036 Nickel	200.8	---	0.002	ND	07/07/06
1042 Potassium	200.7	---	1.0	3.3	07/07/06
1045 Selenium	200.8	0.05	0.008	ND	07/07/06
1050 Silver	200.8	0.1**	0.002	ND	07/07/06
1052 Sodium	200.7	---	1	11	07/07/06
1085 Thallium	200.8	0.002	0.001	ND	07/07/06
1095 Zinc	200.8	5**	0.004	ND	07/07/06

----- This report cannot be reproduced, except in full, without the written approval of National Testing Laboratories, Ltd. -----



NATIONAL TESTING LABORATORIES, LTD.
 556 South Mansfield, Ypsilanti, MI 48197-5166
 (440) 449-2525 • Fax (440) 449-8585



ANALYTICAL REPORT

PAGE 2 OF 2 SAMPLE CODE: 655357

Fed Id #	Analysis Performed	Method	MCL (mg/l)	LRL	Level Detected	Anal Date
Inorganic chemicals - other, and physical factors:						
1927	Alkalinity (Total as CaCO3)	2320B	---	20	64	07/07/06
1004	Bromide		300.0	---	2.0	ND 06/23/06
1017	Chloride		300.0	250**	1.0	98 06/26/06
1910	Corrosivity	Langelier In	---	---	-1.5	07/12/06
1025	Fluoride		300.0	4	0.1	ND 06/23/06
1055	Sulfate		300.0	250**	5.0	6.9 06/23/06
1915	Hardness (as CaCO3)	2340C	---	---	10	180 07/07/06
1930	Total Dissolved Solids	2540C	---	500**	5	260 06/28/06

Fed Id #	Analysis Performed	Method	MCL (mg/l)	LRL	Level Detected	Anal Date	Anal Time
Inorganic chemicals - other, and physical factors:							
1040	Nitrate as N		300.0	10	0.05	1.1	06/23/06 12:47
1041	Nitrite as N		300.0	1	0.05	ND	06/23/06 12:47
1044	Ortho Phosphate		300.0	---	2.0	ND	06/23/06 12:47
1925	pH (Standard Units)		150.1	6.5-8.5**	---	7.0	06/27/06 12:30
4254	pH Temperature (C)					18	
0100	Turbidity (NTU)	2130B	1.0	0.1	ND	06/27/06	14:00
1905	Color	2120B	15**	3.0	ND	06/27/06	13:30
2905	Foaming Agents	5540C	0.5**	0.1	ND	06/27/06	14:40
1920	Odor Threshold	2150B	3 ton**	---	ND	06/27/06	12:10

** Denotes Secondary Maximum Contaminant Level (SMCL)

These test results may be used for compliance purposes as required.

Robert Wegienka

Robert Wegienka, Lab Director